BactoReal® Kit *A. baumannii & P. mirabilis*

Kit version 1.1

Instructions for Use

CE

IVD **For *in vitro* diagnostic use**

REF **DHUS00653**

Σ **50 reactions**

ingenetix GmbH
Haidingergasse 1
1030 Vienna, Austria
T +43(0)1 36 198 01
office@ingenetix.com
www.ingenetix.com

Index

1. Intended purpose 3
2. Product description 3
3. Pathogen information 3
4. Principle of real-time PCR 4
5. Contents of the kit, stability, and storage 4
6. Additionally required materials and devices 5
7. Precautions and safety information 5
7.1. General information 5
7.2. Specific information 5
8. Limitations 6
9. Preparation of samples 7
9.1. Sample collection and storage 7
9.2. Recommended extraction methods 7
9.2.1 Extraction of EDTA blood 7
9.3. Quality control for DNA extraction and PCR inhibition with IPC 7
9.3.1. Application of IPC as control of extraction and real-time PCR 8
9.3.2. Application of IPC as quality control for PCR reaction 8
10. Preparation of real-time PCR 8
10.1. Pipetting scheme 8
10.2. Programming of temperature profile 9
11. Interpretation of data 10
11.1. Controls 10
11.2. Clinical samples 11
12. Troubleshooting 11
12.1. No pathogen specific signal with positive control and with IPC 11
12.2. Pathogen signal with negative control 11
12.3. Pathogen signal with extraction negative control 11
12.4. IPC specific signal with PCR negative control and positive control 12
12.5. No signal with IPC and no pathogen specific signal with sample 12
13. Specifications and performance evaluation 13
13.1. Test performance on different real-time PCR instruments 13
13.2. Limit of detection, LoD95% 13
13.3. Linearity and dynamic range 13
13.4. Precision 14
13.5. Analytical specificity 14
13.6. Clinical evaluation 14
14. Literature 15
15. Revision history 15

NOTICE TO PURCHASER: LIMITED LICENSE

The MGB Probe contained in this product is covered by one or more of the following US patents and corresponding patents outside the US: 5801155 and 6084102 and is sold under a license from the ELITech Group. The purchase of this product includes a license to use only this amount of product solely for the purchaser´s own use solely in the human in vitro diagnostic field (in accordance with applicable FDA and other regulatory requirements) and may not be used for any other commercial use, including without limitation repackaging or resale in any form.

Explanation of symbols

LOT	Batch code		Use by date
REF	Catalogue number		Manufacturer
Σ	Contains sufficient for <n> tests		Temperature limit (Store at)
CE	This product fulfils the requirements of the European Directive 98/79 EC for *in vitro* diagnostic medical devices	IVD	In vitro diagnostic medical device
	Consult instructions for use	UDI	Unique device identifier
	Keep away from sunlight	QTY	Contents

1. Intended purpose

BactoReal® Kit *A. baumannii & P. mirabilis* is a non-automated IVD real-time PCR test for the qualitative detection and identification of DNA of *Acinetobacter baumannii* (rpoB gene) and *Proteus mirabilis* (ureC gene).
Proper specimens are DNA extracts of samples purified from human EDTA blood, aspirates, and CSF.

In combination with other diagnostic tests such as culture this test supports a rapid and specific diagnosis for patients suspected of having a bacterial infection. Results must be interpreted in context of the overall picture and other clinical parameters. The diagnosis of sepsis must not be based solely on BactoReal® Kit *A. baumannii & P. mirabilis.*

The test is intended for professional use and is limited to qualified personnel instructed in the procedures of real-time PCR and *in vitro* diagnostic procedures.

2. Product description

BactoReal® Kit *A. baumannii & P. mirabilis* is a real-time PCR test and detects in the fluorescence channel for FAM the rpoB gene of *Acinetobacter baumannii* and in the fluorescence channel for VIC the ureC gene of *Proteus mirabilis.*

Probe-specific amplification curves in the fluorescence channel for FAM and VIC detect *Acinetobacter baumannii* and *Proteus mirabilis, respectively.* The internal DNA positive control (IPC) is detected in the fluorescence channel Cy5 and serves as a control for DNA extraction and possible real-time PCR inhibition. The target for the DNA IPC (artificial target DNA) is added during sample extraction.

This test has been validated with the ABI® 7500 Fast Real-time PCR System (fast cycle parameters are not supported, Thermo Fisher Scientific) and with the cobas z 480 Analyzer (Roche). However, it is also suitable for other real-time PCR instruments that can measure and differentiate fluorescence in the FAM, VIC and Cy5 channel (e.g., QuantStudio™ 5, QuantStudio™ 7 real-time PCR system (Thermo Fisher Scientific), qTOWER^{3}G (Analytik Jena), Mic instrument (bio molecular systems), LightCycler® 480 II (Roche)).
When using PCR-platforms not tested by ingenetix, an evaluation of the multiplex-PCR shall be done. Keep in mind that some PCR-platforms first must be calibrated with the corresponding dye before performing multiplex-PCR.

Further ingenetix tests for diagnosis of sepsis:
BactoReal® Kit *E. coli & Klebsiella*, order no.: DHUS00153
BactoReal® Kit *Staphylococcus* spp. & *S. aureus*, order no.: DHUS00253
BactoReal® Kit *S. pneumoniae & Streptococcus* spp., order no.: DHUS00353
BactoReal® Kit vanA & *Enterococcus* spp., order no.: DHUS00453
BactoReal® Kit *P. aeruginosa & Enterobacter* spp., order no.: DHUS00553
MycoReal® Kit *Candida & A. fumigatus*, order no.: DHUF00153

Ingenetix BactoReal®, ViroReal®, MycoReal®, PanReal and ParoReal® Kits are optimized to run under the same thermal cycling conditions. DNA and RNA can be analysed in one run.

3. Pathogen information

Acinetobacter baumannii is a gram-negative, aerobic bacterium which is responsible for nosocomial infections. It occurs mainly in intensive care units. This bacterium is often resistant to antibiotics and thus difficult to treat. It can lead to pneumonia, catheter-associated infections and can also cause sepsis.
Proteus mirabilis is a gram-negative, facultatively anaerobic rod-shaped bacterium which frequently occurs in the large intestine and is generally non-pathogenic. However, in people with a weakened immune system, it can lead to urinary tract infections, wound infections, pneumonia, and sepsis.

4. Principle of real-time PCR

The test is based on multiplex real-time PCR by 5´-nuclease-assay technology. Specific DNA sequences are amplified and the generated PCR-products are detected by oligonucleotide-probes labelled with fluorescent dyes. This allows a sequence-specific detection of PCR amplificates.

During PCR, primers are extended by *Taq* polymerase and probes hybridized to the target are cleaved by the 5'-exonuclease activity of *Taq* polymerase. According to the accumulation of PCR product, the fluorescence of the probe increases with each PCR cycle. The change in fluorescence of the different dyes is recorded cycle by cycle in the real-time PCR instrument in the closed reaction tube at different fluorescence wavelengths.
The Cq value (Cq = Quantification cycle, Ct = Cycle threshold, Cp = Crossing point) describes the cycle at which the fluorescence first rises significantly above the background fluorescence.

5. Contents of the kit, stability, and storage

Table 1

Labelling	**Content**	**Amount**	**Storage**
Module 6 + IPC3 Assay Mix (green cap)	Primers and probes for detection of - *Acinetobacter baumannii* (FAM) - *Proteus mirabilis* (VIC) - DNA IPC (Cy5)	1 x 50 µl	-25 to -15 °C
IPC-Target (DNA) (orange cap)	Target for DNA IPC (internal DNA positive control system)	1 x 200 µl	-25 to -15 °C
Module 6 Positive Control (red cap)	DNA positive control (approx. 1,000 target copies/µl)	1 x 300 µl	-25 to -15 °C
DNA Reaction Mix (white cap)	PCR reaction mix for DNA amplification	1 x 500 µl	**-25 to -15 °C after first opening 2 to 8 °C**
Nuclease-free water (blue cap)	Nuclease-free water	1 x 1000 µl	-25 to -15 °C

DNA Reaction Mix
The Master Mix provided with the kit has been designed for reliable, high-sensitivity real-time PCR. The Master Mix contains a highly purified Taq Polymerase for rapid hot-start PCR, dNTPs with dUTP and Uracil-N glycosylase (UNG) to eliminate amplicon carryover, ROX™ dye (passive reference) and buffer components – additives optimized to handle RT-PCR inhibitors.

Delivery and Storage
Shipment is with coolpacks or on dry ice. When stored properly, the kit components are stable until the specified expiration date. This also applies after opening. Store kit protected from light.

Quality Control Release Testing
In accordance with the ISO 13485 certified Quality Management System of ingenetix, each lot is tested against predetermined specifications to ensure consistent product quality. Quality control is performed with a plasmid containing parts of the pathogen DNA.

6. Additionally required materials and devices

- Reagents and devices for DNA-extraction which are appropriate for the listed sample material (see 9. Preparation of the samples)
- Nuclease-free water
- Powder-free disposable gloves
- Pipettes (adjustable)
- Filter pipette tips
- Vortex-Mixer
- Benchtop centrifuge with rotor for 2 ml reaction tubes
- Real-time PCR instrument detecting and differentiating fluorescence in FAM, VIC, Cy5 channel (see 2. Product description)
- Appropriate 96 well optical reaction plates or reaction tubes with corresponding (optical) closing material
- Optional: Laminar flow box
- Optional: PCR workstation

7. Precautions and safety information

7.1. General information

- For *in vitro* diagnostic use. The use of this kit is limited to qualified personnel instructed in real-time PCR and *in vitro* diagnostic procedures.
- Transportation of clinical specimens must comply with local regulations for the transport of biological materials.
- Specimens should be handled as if infectious in accordance with safe laboratory procedures. Wear protective powder-free disposable gloves when handling kit reagents and specimens.
- Discard specimens, reagents, and waste according to your local safety regulations.
- Improper collection, transport or storage of specimens may hinder the ability of the assay to detect the target sequences.
- Quality of DNA has a profound impact on the test performance. Ensure that the used DNA extraction system is compatible with real-time PCR technology.
- The real-time PCR instrument should be calibrated, serviced, and cleaned regularly.
- Protect kit components from light.
- Avoid mixing reagents of different kits and lots and check expiry date of the kit.

7.2. Specific information

A workflow must be followed to avoid false positive results due to detection of contaminating DNA.

<u>Recommended measures to avoid DNA contamination:</u>

- Use separated workspaces for specimen preparation, set-up of real-time PCR and amplification. Supplies and equipment must be dedicated to each of these separate workspaces to ensure workflow in the laboratory from pre- to post-PCR.
- Laboratory benches and devices must be cleaned regularly.
- Sample preparation should be performed in a laminar flow box. Clean laminar flow box regularly in all areas.
- If possible, leave consumables and pipettes in the laminar flow sterile bench and PCR workstation.
- Preparation of real-time PCR should be done in a PCR workstation.
- The use of sterile aerosol-resistant pipette tips is required.
- Use only DNA-free consumables.
- Wear lab coat.
- Work only with powder-free disposable gloves, do not touch the outer surface of the palm and fingers of the gloves when putting them on. Change gloves frequently. To avoid skin contact, wear gloves over the sleeves of the lab coat. Use disposable sleeve protectors if possible.
- Do not touch the rim or threads of open vials.
- Caution is advised when handling specimens and positive control to avoid cross-contamination.
- Store positive and potentially positive material separately from all other reagents.

- For a valid interpretation of results, a negative control shall be included during DNA-extraction (e.g., extraction of water instead of sample material), in to exclude false-positive results due to contamination with DNA of pathogen during extraction.
- Optional: include a negative control of PCR per PCR-run (nuclease-free water instead of sample, NTC).

8. Limitations

- Reliable results with this test are only achieved by appropriate specimen collection, transport, and storage, as well as an appropriate DNA extraction procedure.
- This kit has been used to validate the detection of bacterial DNA from EDTA blood, aspirates, CSF and biopsies. If the concentration of leukocytes in the blood is low, pathogen DNA may be lost during extraction.
- A negative test result does not exclude the possibility of infection with pathogenic organisms.
- Sequence variabilities in the target-region of some subtypes (strains) may lead to false-negative or less sensitive results.
- A pathogen concentration in the sample below the detection limit can lead to a false-negative result.
- Sepsis diagnosis must not be based solely on BactoReal® Kit *A. baumannii & P. mirabilis*. The test has to be used in combination with other diagnostic tests such as culture to enable a rapid and specific diagnosis for patients suspected of having sepsis or a bacterial infection of the bloodstream.
- Results should be interpreted in context of other clinical and laboratory findings.

9. Preparation of samples

BactoReal® Kit *A. baumannii* & *P. mirabilis* is suitable for analysis of DNA extracted from human EDTA blood, aspirates, and CSF.

Sample preparation should be performed with the recommended measures to avoid DNA contamination (see precautions and safety information). Always include a DNA extraction negative control (e.g., extraction of water).

Purified DNA should be stored at -25 to -15 °C.

9.1. Sample collection and storage

Aspirates and cerebrospinal fluid can be stored in microcentrifuge tubes. It is recommended to process samples immediately after collection. Store samples at 2-8 °C for no longer than 48 hours or freeze at -20/-80 °C.

Blood has to be stored in EDTA blood collection tubes. It is recommended to process samples immediately after collection. Store samples at 2-8 °C for no longer than 48 hours or freeze at -20/-80 °C.

9.2. Recommended extraction methods

Ensure that the extraction system used is not contaminated with DNA of pathogens detected with BactoReal® Kit *A. baumannii* & *P. mirabilis.* Extract the sample using a DNA extraction system compatible with real-time PCR technology and appropriate for the sample material.

For manual extraction recommended
- QIAamp DNA Mini Kit (Qiagen)

For automated extraction recommended
- innuPREP AniPath DNA RNA – KFFLX Kit (Analytik Jena) with the KingFisher FLEX instrument (Thermo Fisher Scientific)

9.2.1 Extraction of EDTA blood

The extracted DNA should be equivalent to 50 µl of blood at least (e.g., 500 µl of blood eluted in 100 µl). If the sample is planned to be further screened for other sepsis pathogens, extraction of 1 ml EDTA blood in combination with mechanical lysis by beads is recommended.

For automated extraction of EDTA blood recommended
- Optional: preceding mechanical lysis of 1.3 ml EDTA blood with beads (e.g., MP Biomedicals™ Lysing Matrix E, 2 ml, mpbio) on the Magnalyser or another beadbeater (70 sec, 7000 rpm).
- Subsequently, automated extraction of supernatant (approx. 800 µl) with innuPREP AniPath DNA RNA - KFFLX Kit (Analytik Jena) with a modified protocol for 800 µl sample volume using the KingFisher FLEX extraction device (Thermo Fisher Scientific).

When using extraction methods not recommended by ingenetix, an evaluation of the extraction method must be performed.

9.3. Quality control for DNA extraction and PCR inhibition with IPC

The DNA IPC system (internal DNA positive control) is used as a control for DNA extraction, identifies possible PCR inhibition and confirms the integrity of kit reagents.
For this, an artificial target DNA (IPC-Target (DNA), approx. $6x10^5$ copies/µl) is added during extraction.

Note: The Cq values of the IPC depend on the extraction method and on the type of sample material. Negative samples should show Cq values of the IPC between 27-30. The extraction method used must be validated accordingly with sample material. Use the IPC-Target (DNA) freshly diluted (1:10 with nuclease-free water) in the extraction, if Cq values < 27 are determined during validation with samples.

9.3.1. Application of IPC as control of extraction and real-time PCR

The IPC-Target (DNA) must be added during extraction.

→ Per sample spike 1 µl IPC-Target (DNA) (orange cap) directly into the appropriate volume of lysis buffer (or spike it to the sample after the lysis buffer has been added to the sample), then continue the extraction procedure.

Caution: The IPC-Target (DNA) shall not be added to sample material in the absence of lysis buffer, as degradation may occur. It must be added to the lysis buffer.

9.3.2. Application of IPC as quality control for PCR reaction

If the IPC-Target (DNA) has not been added during extraction, it can be added at a later stage to the PCR master mix.

→ Freshly dilute the IPC-Target (DNA) 1:100 with nuclease-free water and add 1 µl of the dilution/PCR reaction (approx. 6000 target copies).

Caution: The IPC-Target (DNA) shall not be added to the master mix undiluted.

10. Preparation of real-time PCR

- The preparation of the real-time PCR should be done in a PCR workstation.
- Include one positive control (red cap), one extraction negative control and optionally one PCR negative control (NTC, e.g., nuclease-free water) per PCR run.
- It is generally recommended to analyse samples in duplicates, which increases the probability of pathogen detection and facilitates interpretation of results.
- Thaw DNA samples on ice.
- Thaw kit components completely at room temperature. When thawed, mix components carefully, centrifuge briefly with low speed.
- Mix the DNA Reaction Mix gently to ensure homogeneity of solution.
- **Positive Control**
  → Use 9 µl of Positive Control (red cap). Always add Positive Control last.

10.1. Pipetting scheme

		Per sample and module
Preparation Master Mix per module (mix well)	DNA Reaction Mix	10.0 µl
	Module 6 + IPC3 Assay Mix	1.0 µl
	Module Master Mix	**11.0 µl**
PCR-reaction	Module Master Mix	11.0 µl
	DNA-Sample*	9.0 µl
	Total volume	**20.0 µl**

*1-9 µl of the sample can be used. For ≠ 9 µl sample, the volume must be adjusted with nuclease-free water.

→ **If IPC-Target (DNA) was not added during extraction:** Freshly dilute the IPC-Target (DNA) (orange cap) 1:100 with nuclease-free water and add 1 µl per sample directly to the master mix. In this case, the IPC is used for quality control of the PCR reaction. Only 8 µl of DNA sample can be analyzed.
Caution: The use of more than 1 µl diluted IPC-Target (DNA) per reaction might cause inhibition of the real-time PCR reaction.

- Prepare the Master Mix according to the number of samples, calculating an additional volume of approx. 10% for pipetting loss.
- Pipette 11 µl of the prepared Master Mix per sample into the well of the optical reaction plate.
- Then add 9 µl of the extracted sample or controls. Pipette the positive control at last.
- Seal the plate with a suitable optical sealing material.
- Vortex the sealed plate for 1-2 seconds and briefly centrifuge the plate.

10.2. Programming of temperature profile

Please find further information on programming of the real-time PCR instrument in the respective operator's manual. Keep in mind that some PCR-platforms must be calibrated with the corresponding dye before performing a multiplex-PCR.

Sample Volume: 20 µl
Ramp speed: "Standard" without "fast cycling" parameter for ABI® 7500 Fast Instrument, QuantStudio™ (Thermo Fisher Scientific)

Program 1 Cycles: 1 Analysis: None	**Program 2** Cycles: 1 Analysis: None	**Program 3** Cycles: 45 Analysis: Quantification Acquisition at 60 °C
50 °C 2 min	95 °C 2 min	95 °C 5 sec 60 °C 30 sec

Ad program 1: If viral RNA should be also detected in the same PCR run, program 1 must be prolonged to 15 min at 50 °C. This temperature profile can be used for all ingenetix ViroReal®, BactoReal®, MycoReal®, PanReal and ParoReal® kits for the detection of DNA or RNA.
Ad program 2: the previous temperature profile with 20 sec in program 2 can still be used.
Ad program 3: 1 minute possible in combination with other ingenetix kits.

<u>Select detection channels</u>

FAM channel: Detection of *Acinetobacter baumannii*
VIC channel: Detection of *Proteus mirabilis*
Cy5 channel: Detection of IPC

For ABI® 7500 Instrument, QuantStudio™ 5/6/7 (Thermo Fisher Scientific), Mx3005P® (Agilent)
FAM-NONE
VIC-NONE
Cy5-NONE
Passive refence dye: ROX

For cobas z 480 Analyzer / LightCycler® 480 II (Roche)
Detection format: 3 Color Hydrolysis Probe
FAM: Excitation at 465 nm, Emission at 510 nm
VIC: Excitation at 540 / 533 nm, Emission at 580 nm
Cy5: Excitation at 610 / 618 nm, Emission at 670 / 660 nm
Passive reference dye: no ROX needed

The color compensation for FAM and VIC must be selected from the Roche database after the analysis of Cy5.

11. Interpretation of data

For the analysis of the PCR results, select the fluorescence display options FAM and VIC channel for the pathogen target and Cy5 channel for the IPC.
Please note that some PCR platforms require Color Compensation when using a multiplex PCR.

IMPORTANT: Samples should be checked in both logarithmic (Roche instrument: Abs Quant/Fit Points) and linear view and compared to the negative controls. In addition to the Cq values, also check the amplification curves and adjust the threshold (noise band) if necessary.
After you have saved the new settings, export the data. For the cobas z 480 Analyzer, export tables per dye.

Assessment of clinical specimen test results should be performed after the positive and negative controls have been examined and determined to be valid. If results of controls are not valid, no interpretation of results with clinical samples is possible.

Negative samples:
Samples without amplification curves (no Cq values, undetermined) are regarded as negative. No DNA was detected in these samples because there is no infection with the corresponding pathogen or the pathogen DNA concentration is below the detection limit of the test.

Positive samples:
Samples with Cq values < 45 in the fluorescence channel for the pathogen target are regarded as positive.

Table 2 shows the criteria for valid positive and negative controls. Table 3 shows interpretation of data with clinical samples.

11.1. Controls

Table 2 Criteria for valid controls, IPC-Target (DNA) was added during extraction.

	Cq FAM channel *A. baumannii*	Cq VIC channel *P. mirabilis*	Cq Cy5 channel IPC[1]	Interpretation	Action
Positive control	<31	<31	Negative	Valid	-
Positive control	Negative	Negative	Negative	Invalid	See 12.1
Positive control	<31	Negative	Negative	Invalid	See 12.1
Positive control	Negative	<31	Negative	Invalid	See 12.1
Extraction negative control	Negative	Negative	27-30[2]	Valid	-
Extraction negative control	Negative	Negative	Negative	Invalid	See 12.1
Extraction negative control	Positive	Positive	27-30[2]	Invalid	See 12.3
Extraction negative control	Positive	Negative	27-30[2]	Invalid	See 12.3
Extraction negative control	Negative	Positive	27-30[2]	Invalid	See 12.3
NTC PCR (optional)	Negative	Negative	Negative	Valid	-
NTC PCR (optional)	Positive	Positive	Negative	Invalid	See 12.2
NTC PCR (optional)	Negative	Positive	Negative	Invalid	See 12.2
NTC PCR (optional)	Positive	Negative	Negative	Invalid	See 12.2

[1] If the IPC-Target (DNA) was added directly to the master mix, all samples in the Cy5 channel must be positive.
[2] Cq values of the IPC are dependent on the extraction method and should show Cq values of 27-30 (see 9. preparation of samples).

11.2. Clinical samples

Samples with positive Cq values are interpreted as positive, see Table 3.

Table 3 Interpretation of clinical samples

	Cq FAM channel *A. baumannii*	Cq VIC channel *P. mirabilis*	Cq Cy5 channel IPC	Interpretation	Action
Clinical sample	Negative	Negative	27-30[1]	Negative	-
Clinical sample	Positive	Positive	Positive/Negative[2]	Positive for *A. baumannii* + *P. mirabilis*	-
Clinical sample	Negative	Positive	Positive/Negative[2]	Positive for *P. mirabilis*	-
Clinical sample	Positive	Negative	Positive/Negative[2]	Positive for *A. baumannii*	-
Clinical sample	Negative	Negative	Negative	Invalid	See 12.5

[1] A positive signal of the IPC excludes potential PCR inhibition. However, IPC Cq-values should show comparable results among samples. A shift of Cq- values can indicate a partial inhibition of PCR. Cq values of the IPC are dependent on the extraction method and should show Cq values of 27-30 (see 9. preparation of samples).
[2] High pathogen loads in the sample can lead to a reduced or absent fluorescence signal of the IPC.

In case of invalid data, the analysis must be repeated with the remaining or newly extracted DNA sample (see 12. Troubleshooting).

12. Troubleshooting

12.1. No pathogen specific signal with positive control and with IPC

- Incorrect programming of the temperature profile or incorrect setting of detection channels on the real-time PCR instrument.
  → Compare temperature profile and setting of detection channels with details specified in the protocol
- Incorrect configuration of PCR reaction.
  → Check your pipetting steps with the pipetting scheme and repeat PCR, if necessary.
  → The DNA may have been degraded.
- The IPC-Target (DNA) was added undiluted directly to the master mix and not freshly diluted 1:100. The PCR reaction is therefore inhibited.
  → Freshly dilute IPC-Target (DNA) 1:100 and repeat PCR.
- No Positive Control was added.
  → Repeat PCR in case all clinical samples are negative.
- For control of real-time PCR only: 1 µl of freshly 1:100 diluted IPC-Target (DNA) must be added to the master mix. If the addition of IPC-Target has been forgotten:
  → Freshly dilute IPC-Target (DNA) and repeat PCR.
- For control of DNA extraction and PCR inhibition, the undiluted IPC-Target (DNA) must be added during extraction to the lysis buffer. If the addition of IPC-Target has been forgotten:
  → Repeat DNA extraction.

12.2. Pathogen signal with negative control

- If there are Cq values in the negative control below the cut-off specified in the Instructions for Use, a contamination occurred during the preparation of the PCR.
  → Repeat PCR with new reagents in replicates.
  → Strictly pipette positive control at last.
  → Make sure that workspace and instruments are cleaned at regular intervals.

12.3. Pathogen signal with extraction negative control

- If there are Cq values in the negative control of extraction below the cut-off specified in the Instructions for Use, a contamination occurred during the preparation of the PCR or the DNA extraction.

→ Repeat the PCR with unused reagents in replicates.
→ If repeatedly positive: Repeat DNA extraction and PCR using new reagents.
→ Strictly pipette positive control at last.
→ Make sure that workspace and instruments are cleaned at regular intervals and that workstations are separated from samples with high pathogen concentration.

12.4. IPC specific signal with PCR negative control and positive control

- The IPC-Target (DNA) has been added to lysis buffer during extraction, but there is IPC specific signal with PCR-NTC and positive control: Contamination with the IPC-Target (DNA).
  → Make sure that workspace and instruments are cleaned at regular intervals.

12.5. No signal with IPC and no pathogen specific signal with sample

- Incorrect assignment of detection channels in sample.
  → Please verify the correct assignment of detection channels.
- The DNA might be degraded.
- If the IPC-Target (DNA) was added during extraction:
  - Inhibition of PCR may have occurred.
  - DNA extraction was unsuccessful.
  - The IPC-Target (DNA) was not added to the lysis buffer of the sample.
  - The extracted sample was not added to the PCR-reaction.

  → No statement is possible. Verify you use a recommended method for DNA isolation and re-examine the single steps of the DNA extraction.
  → If no operating mistakes during DNA extractions can be retraced, it is recommended to repeat the PCR with lower amounts of DNA-eluate (1/5 or 1/10 of sample volume + the adequate volume of nuclease-free water).

13. Specifications and performance evaluation

13.1. Test performance on different real-time PCR instruments

Performance of BactoReal® Kit *A. baumannii* & *P. mirabilis* with an ABI® 7500 instrument (Thermo Fisher Scientific) is shown in Figure 1.

Figure 1 Performance of BactoReal® Kit *A. baumannii & P. mirabilis*



Positive Control: FAM (*A. baumannii*), VIC (*P. mirabilis*) and CY5 channel

This test has been validated with the ABI® 7500 instrument (Thermo Fisher Scientific) and with the cobas z 480 Analyzer (Roche).

13.2. Limit of detection, LoD95%

Method: BactoReal® Kit *A. baumannii* & *P. mirabilis* was tested with dilution series of plasmids containing parts of the respective target DNA. The limit of detection (LoD95% = smallest number of copies of target DNA which can be detected in 95% of cases) was determined with plasmid dilutions around the analytical sensitivity using a non-linear curve fitting with the Graph Pad Prism Software.

Results: The results are shown in Table 4.
Note: Sequence variability in the primer or probe region may lead to false-negative or less sensitive results for some pathogen strains.

Table 4 Summary of LoD95% and linearity

	LoD95% Without / with cut-off rule	Linearity – Slope	Linearity – R^2
A. *baumannii*	7	-3.26 ± 0.04	0.9936
P. *mirabilis*	7	-3.37 ± 0.04	0.9930

13.3. Linearity and dynamic range

Method: Linearity was determined using 10-fold dilution series (10^6 – 10^1 target copies/reaction) of the plasmids. The number of determinations (n) per dilution was nine.

Results: Results are shown in Table 4. The test shows linearity over the range of 100 to 1,000,000 target copies/reaction.

13.4. Precision

Method: Precision within a run (intra-assay), between multiple runs (inter-assay) and between two lots (inter-lot) was determined.

Results: The mean values of the coefficients of variation (CV%) are summarized in Table 5.

Table 5 Summary of mean values of the coefficients of variation

	Coefficients of variation intra-assay	Coefficients of variation inter-assay	Coefficients of variation inter-lot
A. *baumannii*	0.84	0.54	0.51
P. *mirabilis*	0.76	0.88	1.80

13.5. Analytical specificity

Method: The selection of highly specific primers and probes ensures analytical specificity. The specificity of primer and probes was validated *in silico* using the Basic Local Alignment Search Tool (BLAST) against the NCBI database. Primers and probes have been checked for possible homologies to currently published sequences. This database analysis validated the detection of currently known strains and potential cross-reactions.
Specificity was further determined by testing different bacterial and fungal isolates.

Results: The results of the BLAST analyses in terms of detection of different species are summarized in Table 6. Sequence variability in the target region of some subtypes (strains) may lead to false-negative or less sensitive results.
When testing different bacterial and fungal isolates, no significant cross-reactions were observed (data not shown).

Table 6 Results of BLAST analysis

	Detected species
FAM channel, A. *baumannii*	*Acinetobacter baumannii*
VIC channel, P. *mirabilis*	*Proteus mirabilis*

13.6. Clinical evaluation

Method: 7 blood samples (extracted with SeptiFast LysKit, Roche) and 5 other specimens such as CSF and aspirates of patients with suspected bacterial infection were analyzed. The results were compared with the LightCycler® SeptiFast Test (Roche) for blood samples and the SepsiTest™ (Molzym) for the other test materials

Results: see Table 7.

Table 7 Statistical clinical evaluation for blood samples and other samples (CSF, aspirates)

	Sensitivity		Specificity		NPV		PPV		Prevalence		Accuracy	
	Blood	Other	Blood	Other	Blood	Other	Blood	Other	Blood	Other	Blood	Other
A. baumannii	100.0%	100.0%	100.0%	-	100.0%	100.0%	100.0%	-	14.3%	0.00%	100.0%	-
P. mirabilis	100.0%	100.0%	100.0%	-	100.0%	100.0%	100.0%	-	14.3%	0.00%	100.0%	-

14. Literature

1. Angus DC, Wax RS. Epidemiology of sepsis: an update. Critical care medicine. 2001; 29 (7 Suppl):S109.
2. Hotchkiss R S, Karl IE. The pathophysiology and treatment of sepsis. N Engl J Med. 2003; 348:138.
3. Makristathis A, Riss S, Hirschl AM. A novel fluorescence in situ hybridization test for rapid pathogen identification in positive blood cultures. Clin Microbiol Infect. 2014;20:760.
4. Mitteregger D, Barousch W, Nehr M, Kundi M, Zeitlinger M, Makristathis A, Hirschl AM. Neutralization of antimicrobial substances in new BacT/Alert FA and FN Plus blood culture bottles. J Clin Microbiol. 2013;51:1534.
5. Westh H, Lisby G, Breysse F, Böddinghaus B, Chomarat M, Gant V, Goglio A, Raglio A, Schuster H, Stuber F, Wissing H, Hoeft A. Multiplex real-time PCR and blood culture for identification of bloodstream pathogens in patients with suspected sepsis. Clin Microbiol Infect. 2009;15:544.
6. Lehmann LE, Hunfeld KP, Emrich T, Haberhausen G, Wissing H, Hoeft A, Stüber F. A multiplex real-time PCR assay for rapid detection and differentiation of 25 bacterial and fungal pathogens from whole blood samples. Med Microbiol Immunol. 2008;197:313.

15. Revision history

Revision	Date	Description
1.1en	01 May 2023	**Change of the company address** **1. Intended use** Intended use was specified in more detail. **5. Contents, stability and storage** Kit version 1.1: Filling volume of IPC-Target (DNA) and Positive Control was changed. The DNA IPC Target is not stored in RNA/DNA Stabilizer anymore and was renamed to IPC-Target (DNA). Module 6 Assay Mix was renamed to Module 6 + IPC3 Assay Mix. **6. Additionally required materials and equipment** Update **7. Precautions and safety instructions** Update of this section Inclusion of additional information on specimen collection and storage, processing of specimens after collection, storage conditions of specimens, transport and storage conditions of purified DNA. Recommended measures to avoid DNA contamination added. **9. Preparation of samples** Information on sample storage was added. MagMAX™ Viral/Pathogen Ultra Nucleic Acid Isolation Kit has been removed. innuPREP AniPath DNA RNA - KFFLX Kit added Note that for Cq values < 27, the IPC-Target (DNA) must be diluted prior to extraction. **10. preparation of real-time PCR** The temperature profile has changed - the activation step in program 2 has been extended from 20 sec to 2 min. **11. Interpretation of PCR data** Definition of Cq, Ct and Cp values. Update of tables for results with positive and negative controls and clinical samples. Inclusion of an explanation of how to evaluate clinical specimen test results in relation to controls. **12. Troubleshooting** Update regarding references in the new results tables. **13. Specifications and performance evaluation** Update

Note:
Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

Technical support
ingenetix GmbH, Haidingergasse 1, 1030 Vienna, Austria
Telephone: +43 (0)1 36 198 01; **E-Mail:** office@ingenetix.com