

ViroReal[®] Kit PEDV

Manual

For use with the

- ABI PRISM[®] 7500 (Fast)
- LightCycler[®] 480
- Mx3005P[®]



For veterinary use only



DVEV01313



100



DVEV01353



50



ingenetix GmbH
 Arsenalstraße 11
 1030 Vienna, Austria
 T +43(0)1 36 198 0 198
 F +43(0)1 36 198 0 199
 office@ingenetix.com
 www.ingenetix.com

Index

1. Product description	3
2. Pathogen information	3
3. Principle of real-time PCR	3
4. General Precautions	3
5. Contents of the Kit.....	4
6. Additionally required materials and devices.....	4
7. Preparation of real-time PCR	5
7.1. Internal RNA positive control	5
7.2. Positive Control.....	5
7.3. Pipetting scheme	5
7.4. Programming of the temperature profile	6
8. Interpretation of PCR-data	6
9. Troubleshooting.....	8
9.1. No PEDV specific signal with positive control:	8
9.2. No signal with internal RNA positive control and no PEDV specific signals with sample:	8
9.3. PEDV specific signal with negative control:.....	8
9.4. PEDV specific signal with negative control of RNA-extraction (optional):	8
10. Specifications	8
10.1. Analytical sensitivity	8
10.2. Analytical specificity	8
11. Annex – symbols.....	8

1. Product description

ViroReal® Kit PEDV is a real-time PCR assay for detection of RNA of the porcine epidemic diarrhea virus (PEDV) using one-step reverse transcription real-time PCR. This test was developed and validated for the ABI PRISM® 7500 (Fast) instrument (Thermo Fisher Scientific), LightCycler® 480 (Roche) and Mx3005P® (Agilent), but is also suitable for other real-time PCR instruments.

ViroReal® Kit PEDV detects the nucleocapsid protein gene (N gene) of PEDV. A probe-specific amplification-curve in the FAM channel indicates the amplification of PEDV RNA. This kit allows the rapid and sensitive detection of PEDV RNA isolated from feces of acutely affected pigs. Virus RNA can be recovered efficiently from sample material using for example the QIAamp Viral RNA Mini Kit (Qiagen) extraction methods.

An internal RNA positive control system for detection in Cy5 channel allows control of RNA extraction and excludes false-negative interpretation of results due to inhibition of reverse transcription real-time PCR (see 8. Interpretation of PCR-data).

When using PCR-platforms not validated by ingenetix, an evaluation of the multiplex-PCR is recommended. Please be aware that some PCR-platforms have to be calibrated with the corresponding dye before performing multiplex-PCR.

BactoReal®, MycoReal, ParoReal and ViroReal® Kits are optimized to run under the same thermal cycling conditions. RNA and DNA material can be analysed in one run.

2. Pathogen information

Porcine epidemic diarrhea virus (PEDV), a member of the genus *Coronavirus (Alphacoronavirus)*, family *Coronaviridae*, is a positive-sense, enveloped, single-stranded RNA virus. Severity of disease is variable and dependent on the epidemiological status of the herd. When epidemic, PEDV causes acute watery diarrhea and vomiting in a large proportion at all ages of swine. If endemic, then diarrhea is observed with lower morbidity in suckling and recently weaned pigs. The PED virus is similar to, but antigenically distinct from transmissible gastroenteritis virus (TGEV).

References:

Kim O et al. 2002. Comparison of reverse transcription polymerase chain reaction, immunohistochemistry, and in situ hybridization for the detection of porcine epidemic diarrhea virus. *Can J Vet Res.* 66:112-116.

3. Principle of real-time PCR

When detecting pathogens by reverse transcription real-time PCR, a specific RNA sequence of the pathogen genome is transcribed into cDNA and amplified. The generated PCR-product is detected by an oligonucleotide-probe labelled with a fluorescent dye. This technology allows for a sequence-specific detection of PCR amplicates.

4. General Precautions

The user should always pay attention to the following:

- Always include a negative control per PCR-run (Nuclease-free water instead of sample).
- Optional: for valid interpretation of results, a negative control should be included during RNA-extraction (for example extraction of water instead of sample material), in order to exclude false-positive results due to contamination with PEDV RNA during extraction.
- Be careful when handling the positive control.
- Store and extract positive material (specimens, controls and amplicons) separately from all other reagents and add it to the reaction mix in a spatially separated workspace.
- Periodically decontaminate benches and devices.
- Use sterile pipette tips with filters.
- Thaw all components thoroughly at room temperature before starting an assay. When thawed, mix the components and centrifuge briefly.
- Always keep the RNA Reaction Mix on ice.
- Use the RNA immediately after extraction and store at -20°C to -80°C as soon as possible.
- **Caution:** the RNA IPC Target is stored in RNA stabilizer which contains guanidinium thiocyanate/Triton X-100 (see MSDS, www.ingenetix.com).

5. Contents of the Kit

Labelling	Content	Amount		Storage
		DVEV01313	DVEV01353	
PEDV Assay Mix (green cap)	Primer and probe (FAM) for PEDV detection	2 x 50 µl	1 x 50 µl	-20°C
RNA IPC-3 Assay Mix (yellow cap)	Primer and probe (Cy5) for RNA IPC detection	2 x 50 µl	1 x 50 µl	-20°C
RNA IPC Target (orange cap)	RNA internal positive control	1 x 100 µl	1 x 100 µl	-20°C
PEDV Positive Control (red cap)	RNA positive control	1 x 300 µl	1 x 300 µl	-20°C
RNA Reaction Mix (white cap)	4 x Reaction Mix	2 x 250 µl	1 x 250 µl	-20°C
Nuclease-free water (blue cap)	Nuclease-free water	2 x 1000 µl	1 x 1000 µl	-20°C

The components of ViroReal® Kit PEDV are stable until the expiry date stated on the label. Repeated thawing and freezing should be avoided. Please protect kit components from light.

6. Additionally required materials and devices

- Reagents and devices for RNA-extraction
- Nuclease-free water for dilution of RNA IPC Target
- Disposable powder-free gloves
- Pipettes (adjustable)
- Sterile pipette tips with filters
- Vortex mixer
- Desktop centrifuge with rotor for 2 ml reaction tubes
- Real-time PCR instrument which is able to detect and differentiate fluorescence in FAM and VIC/HEX or Cy5 channel
- Appropriate 96 well reaction plates or reaction tubes with corresponding (optical) closing material

7. Preparation of real-time PCR

Please make sure that at least one negative control (water, blue cap), as well as one positive control (red cap) and one extraction negative control (optional, recommended) are included per PCR run.

Ingenetix highly recommends performing PCR analyses in duplicates, which increases the probability of detection of the pathogen and facilitates interpretation of results.

- Prepare master mix on ice.
- Thaw RNA Reaction Mix on ice, and invert 2 to 3 times to ensure homogenous solution. Do not let it warm to room temperature.
- Use RNA immediately after extraction and store at -20 to -80°C as soon as possible.

7.1. Internal RNA positive control

An internal RNA positive control system containing the RNA IPC assay and the RNA IPC Target excludes false-negative interpretation of results due to inhibition of reverse transcription real-time PCR.

→ Dilute RNA IPC Target freshly 1:500 with nuclease-free water and add to the master mix (use 1 µl/reaction).

→ Alternatively, for control of RNA extraction and PCR inhibition the RNA IPC Target can be added during extraction. Spike 1 µl of undiluted RNA IPC Target into the sample material after the lysis buffer was added.

Caution: Do not add the RNA IPC Target directly to the sample material.

Optional: 1:10 dilution of the 1:500 diluted positive control can be used and defined as second standard value (approx. 10,000 target copies/µl).

7.2. Positive Control

The PEDV Positive Control is an *in vitro* synthesized RNA sample with a concentration of approx. 10³ copies/µl. It has to be stored at -20°C. Ensure a homogenous solution by gently mixing, do not vortex. To avoid freeze/thaw cycles, it can also be temporarily stored at 4°C if used several times on the same day.

→ As positive control please use 10 µl of the Positive Control.

7.3. Pipetting scheme

		Per sample
Preparation of Master Mix (mix well)	Nuclease-free Water*	2.0 µl
	RNA Reaction Mix	5.0 µl
	PEDV Assay Mix	1.0 µl
	RNA IPC Assay Mix	1.0 µl
	RNA IPC Target [#] (freshly diluted 1:500)	1.0 µl
	Total volume Master Mix	10.0 µl
Preparation of PCR	Master Mix	10.0 µl
	RNA-Sample*	10.0 µl
	Total volume	20.0 µl

*1-10 µl of the sample can be used. When using an amount < 10 µl of the sample, the amount of H₂O has to be changed accordingly.

[#]If RNA IPC Target not already added during extraction.

7.4. Programming of the temperature profile

Please find further information on programming the real-time PCR instrument in the respective operator's manual. Please be aware that some PCR-platforms have to be calibrated with the corresponding dye before performing multiplex-PCR.

Select dyes: FAM-TAMRA (530 nm) for detection of PEDV
 Cy5-NONE (RNA IPC-3 Assay Mix) for detection of RNA IPC

Select reference dye (passive reference): ROX

Sample Volume: 20 µl

Temperature Profile:

Program 1	Program 2	Program 3
Cycles: 1 Analysis: None	Cycles: 1 Analysis: None	Cycles: 45 Analysis: Quantification Acquisition at 60°
50°C 15 min	95°C 20 sec	95°C 5 sec 60°C 1 min

For ABI PRISM® 7500:
 Ramp speed: Without “fast cycling” parameter

For LightCycler® 480 instrument:
 Detection format: 2 Color Hydrolysis Probe
 (dyes see above)

Note: These instrument parameters can be used for all BactoReal®, MycoReal, ParoReal and ViroReal® kits on all PCR instruments.

8. Interpretation of PCR-data

Examples for interpretation of positive reactions are shown in the amplification plots below.

For a valid interpretation, the following criteria must be fulfilled:

	Ct/Cp (FAM channel) PEDV target	Ct/Cp (Cy5 channel) RNA IPC target	Interpretation
Negative control	Negative	26-29*	Valid
Positive control	<30	26-29	Valid
Extraction negative control (optional)	Negative	26-29	Valid
Negative sample	Negative	26-29	Valid
Positive sample	Positive	26-29/negative	Valid

*In the case that the RNA IPC target has been added to the master mix

For analysis of PCR data please proceed as follows:

For analysis of PCR results gained with ViroReal® Kit PEDV please select fluorescence display options 530 nm (FAM channel) for the PEDV target and Cy5 channel for the RNA IPC target. Samples with a positive Ct or Cp-value are considered positive. Please also check the presence of amplification-curves manually.

8.1. Signal in FAM channel:

→ RNA of PEDV was amplified. The sample has to be interpreted as positive.

8.2. No signal in FAM channel but signal of the internal RNA positive control:


→ No RNA of PEDV is detectable in the sample. The sample has to be interpreted as negative. The positive signal of the internal positive control assay excludes a putative PCR inhibition.

8.3. No signal in Cy5 channel:

→ No interpretation statement can be made.


Information about possible sources of error and their solution can be found in 9. Troubleshooting.

Detection of PEDV




ABI Prism® 7500: FAM channel, 530 nm
1:10 serial dilution of a PEDV positive control


Detection of internal RNA positive control




ABI Prism® 7500: Cy5 channel, 667 nm
Detection of internal RNA positive control




Mx3005P®: FAM channel
1:10 serial dilution of a PEDV positive control



Mx3005P®: CY5 channel
Detection of internal RNA positive control



LightCycler® 480: FAM channel
1:10 serial dilution of a PEDV positive control



LightCycler® 480: Cy5 channel
Detection of internal RNA positive control

9. Troubleshooting

9.1. No PEDV specific signal with positive control:

- Incorrect programming of the temperature profile of the real-time PCR instrument.
→ Compare the temperature profile with the protocol (see 7. Preparation of real-time PCR).
- Incorrect configuration of the PCR reaction.
→ Check your work steps (see 7. Preparation of real-time PCR) and repeat the PCR, if necessary.
- RNA might be degraded.

9.2. No signal with internal RNA positive control and no PEDV specific signals with sample:

- The PCR reaction was inhibited. No interpretation can be made.
→ Make sure that you use a recommended method for RNA isolation and stick closely to the manufacturer's instructions.
→ If no operating mistakes during extractions can be retraced, it is recommended to repeat the PCR with lower amounts of RNA-eluate (1/5 or 1/10 of sample volume + the adequate amount of H₂O).
- Incorrect PCR conditions.
→ Check the PCR conditions and repeat the PCR, if necessary.

9.3. PEDV specific signal with negative control:

- A contamination occurred during preparation of the PCR.
→ Repeat PCR with new reagents in replicates.
→ Strictly pipette the positive controls at last.
→ Make sure that work space and instruments are decontaminated at regular intervals.

9.4. PEDV specific signal with negative control of RNA-extraction (optional):

- A contamination occurred during extraction.
→ Repeat the extraction and PCR using new reagents.
→ Make sure that work space and instruments are decontaminated at regular intervals.

10. Specifications

ViroReal[®] Kit PEDV was evaluated with the ABI PRISM[®] 7500 (Fast) instrument (Thermo Fisher Scientific), with the LightCycler[®] 480 (Roche) and the Mx3005P[®] (Agilent). For further validation data please contact ingenetix.

10.1. Analytical sensitivity

The analytical sensitivity is 10 RNA copies/PCR. The limit of detection (LoD₉₅ = smallest number of copies of target RNA which can be detected in 95% of cases) is 37 target copies/reaction and was determined by several replicates around the detection limit.

10.2. Analytical specificity

The specificity is ensured by the selection of highly specific primers and probes. The primers and probes were checked for possible homologies to currently published sequences by sequence comparison analyses. This also validated the detection of so far known PEDV strains published in the NCBI database. The kit was tested with one TGEV, one PHEV, three PEDV, eight PRRSV EU, four PPV and seven PCV2 strains. It was positive with PEDV and showed no cross-reaction with the others.

11. Annex – symbols



Batch code



Catalogue number



Contains sufficient for <n> tests



Corrosion, GHS05



Use by



Manufactured by



Store at



Exclamation mark, GHS07